



# Gauhati University

Department of Chemistry  
Gopinath Bardoloi Nagar, Guwahati-781014  
Assam::India

NIQ No: GU/Chem/DST-INSPIRE/TKG/2016-2

Date: 21/12/2016  
Place: Guwahati

Quotations from the reputed manufacturers/dealers/suppliers are invited for the supply, installation etc. for an **Electronic Balance** in the department of Chemistry, Gauhati University, Guwahati, Assam under the DST-INSPIRE Faculty Program (File No. IFA14-CH-164). **The last date of submitting the quotation to the undersigned is 31/12/2016.**

## Item: Precision Laboratory Balance (1 unit)

Technical Specifications: (i) Capacity: ~200g; (ii) Readability: 0.001g; (iii) Repeatability: 0.001g; (iv) Linearity: 0.002g; (v) Pan Size: 90mm; (vi) Tare Range: Full; (vii) Operating Temperature: 10-40°C; (viii) Power Supply: 230V.

**The vendor must have ISO certificate and service center in Guwahati, Assam.**

Gauhati University reserves the right to modify/cancel the requirements.

All terms and conditions will be as per the Gauhati University rules, regulations and decisions.

## Principal Investigator:

*Tridib Kumar Goswami*

Dr. Tridib Kumar Goswami  
Assistant Professor  
Department of Chemistry  
Gauhati University, Guwahati-781014, Assam  
E-mail: tridibgoswami05@gmail.com  
Phone: 8761849858

Dr. Tridib Kumar Goswami  
Assistant Professor  
Department of Chemistry  
Gauhati University