



**GAUHATIUNIVERSITY**  
**Department of Instrumentation & USIC**  
**GopinathBardoloi Nagar, Guwahati – 781 014 Assam : India**  
Contact: +919435545453 (M), *email: utpalsarma@gauhati.ac.in*

**Ref.: INSTR&USIC/NIQ/18/11**

**Date: 08.01.2018**

From: Dr. UtpalSarma  
Professor & Head

Notice Inviting Quotation for Supplying Laboratory Equipment

Sealed quotations from the registered dealers/suppliers are invited for supplying the items in the annexed sheet with mentioned specification to the Department of Instrumentation & USIC, G.U. The quotations must be submitted to the HOD; Department of Instrumentation & USIC, G.U. within 10 days from the date of notification of this NIQ in G.U. website. GST; Unit Price; Model No. HSN Code etc should be shown clearly in the quotations.

**Terms and Condition:**

1. Gauhati University (GU) reserves the right to modify/cancel the requirements without any further notice.
2. All quotations will be treated as per GU rules and regulations.
3. Quoted rates **MUST** be valid for at least 90 days.
5. Lowest bid will not be the only criterion for selection of dealers/suppliers.

(Utpal Sarma)

Copy to :-

1. System Administrator, GU, with a request to upload in GU Website
2. Department Notice Board.

## List of Items

Ref.: INSTR&amp;USIC/NIQ/18/

Date: 08.01.2018

SI No	Item	Specification
1	Oscilloscope Probe	Bandwidth :150MHz Input C :8.5~18.5pF Max input Vol.:600V Dc (1:10) AC Peak (1:10) <b>Warranty: One Year</b>
2	DIGITAL MULTIMETER – True RMS, auto ranging	Display: 3½ digit, 6,000 counts, auto/manual range Dc Voltage: 60mV - 1000V ( (0.1% ) AC Voltage: 60mV to 750V DC Amp: 600µA to 10A in 6 ranges AC Amp : 600µA to 10A in 6 ranges Resistance: 600 Ω to 60M.Ω Capacitance: 60nF to 4000µF in 6 ranges Frequency: 10Hz to 10MHz Other Features: Analog Bar Graph, True RMS, Continuity, diode test, Duty Cycle, Data Hold, Peak Hold, RS-232C interface. Back Light, Auto Power Off facility. Accessory: Battery, spare fuse, test leads, manual, PC Software. <b>Warranty: 5 Years</b>

(Utpal Sarma)